

DEC 03 2010

**VERIFICATION OF CALIBRATION REPORT**  
*of DataMaster cdm Breath Test Instrument*  
*State of Alaska*  
*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

<b>DataMaster cdm S/N:</b> <u>130113</u>	
Supervisor/Operator Performing the Verification Procedure:	
Name: <u>Kenneth G. Cox</u> ID #: <u>3527</u> Date: <u>December 27, 2009</u>	
A	Agency: <u>Skagway Police Department</u> Phone #: <u>907 983-2232</u>
Instrument Location <u>Skagway Police Department, 1<sup>st</sup> &amp; State Street Skagway, Alaska 99840</u>	
Alco S/N	<u>X173010</u> Target Value <u>.082</u> High Pressure <u>300</u>
Alco Test Value Average	<u>.083</u> <u>.083</u>
	1 <sup>st</sup> Alco 2 <sup>nd</sup> Alco
Signature <u>[Signature]</u>	
(OVER)	

(Do Not write in the area below)


[Signature]  
1/9/10

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 26<sup>th</sup> day of January, 2010.  
[Signature] (Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130113

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130113

DECEMBER 27, 2009

OPERATOR'S NAME:  
COX/KENNETH/G  
OPERATOR'S NUMBER: 3527  
SUBJECT'S LAST NAME:  
VERIFICATION OF CAL  
SUBJECT'S FIRST NAME/MI :  
L

O.L. #: L  
DEPT/AGENCY: SK61  
CASE/REPORT: 2009-4594  
TEST TYPE: V  
ALCO TARGET VALUE: .082  
ALCO S/N: X173010

### BREATH ANALYSIS

.082 ADJUSTED FOR 30.32 in  
ALCO TARGET .083 15:49  
BLANK TEST .000 15:50  
INTERNAL STANDARD VERIFIED 15:50  
ALCO TO 30.32 in .083 15:51  
BLANK TEST .000 15:52  
SUBJECT SAMPLE .000 15:52  
BLANK TEST .000 15:53  
ALCO TO 30.32 in .083 15:54  
BLANK TEST .000

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130113

DECEMBER 27, 2009  
TIME 15:43

### DIAGNOSTIC CHECK

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 41c  
BAROMETER: 30.32 in.  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

!@#%&'()\*+,-./0123456789:;<=>?@ABCDEF6  
HIJKLMNPOQRSTUVWXYZ[\]^\_`abcdefgijklmno  
pqrs tuvwxyz{|}~